



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

A & D INSTRUMENTS INDIA PRIVATE LIMITED (CALIBRATION DIVISION), PLOT NO. D-48, UDYOG VIHAR, PHASE-V, GURUGRAM, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3722

Page No

1 of 2

Validity

05/10/2023 to 04/10/2025

Last Amended on

07/11/2023

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
Permanent Facility					
1	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Balance Accuracy Class I and Coarser (Readability: 0.01 mg)	Using E1 Accuracy Class Weights based on OIML R-76-1 by Comparison Method	0 to 250 g	0.123mg
2	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Balance Accuracy Class I and Coarser (Readability: 0.001 mg)	Using E1 Accuracy Class Weights based on OIML R-76-1 by Comparison Method	0 to 22 g	0.02mg
3	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Balance Accuracy Class I and Coarser (Readability: 0.1 mg)	Using E1 Accuracy Class Weights based on OIML R-76-1 by Comparison Method	0 to 520 g	0.25mg
4	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Balance Accuracy Class II and Coarser (Readability: 0.01 g)	Using E1 Accuracy Class Weights based on OIML R-76-1 by Comparison Method	0 to 10200 g	24mg
5	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Balance Accuracy Class II and Coarser (Readability: 0.1 g)	Using E1 Accuracy Class Weights based on OIML R-76-1 by Comparison Method	0 to 61000 g	0.4g
6	MECHANICAL-WEIGHING SCALE AND BALANCE	Electronic Balance Accuracy Class II and Coarser (Readability: 1 mg)	Using E1 Accuracy Class Weights based on OIML R-76-1 by Comparison Method	0 to 1620 g	2mg



National Accreditation Board for
Testing and Calibration Laboratories

CERTIFICATE OF ACCREDITATION

**A & D INSTRUMENTS INDIA PRIVATE LIMITED
(CALIBRATION DIVISION)**

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

PLOT NO. D-48, UDYOG VIHAR, PHASE-V, GURUGRAM, HARYANA, INDIA

in the field of

CALIBRATION

Certificate Number: CC-3722

Issue Date: 05/10/2023

Valid Until:

04/10/2025

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Entity: A & D INSTRUMENTS INDIA PRIVATE LIMITED

Signed for and on behalf of NABL



N. Venkateswaran
Chief Executive Officer